

Appendix C.

April 04, 2007 Email from Bennett Sample Pumps, Inc. informing Citizen Action New Mexico that Bennett groundwater sampling pumps can be modified with a flow valve for low-flow purging and sampling of the monitoring wells at the Sandia Mixed Waste Landfill (Sandia MWL dump).

-----Original Message-----

From: Bennett Sample Pumps, Inc. [mailto:bspinc@arn.net]

Sent: Wednesday, April 04, 2007 3:49 PM

To: dave@radfreenm.org

Subject: Low flow with Bennett Pump

Dave McCoy:

Flow rates lower than those indicated on our flow curves can be achieved by installing a valve in the water discharge line of the tube bundle if this is an acceptable sampling procedure for the customer. When the flow curves were developed the primary interest was in higher flow rates. With the valve, the flow rate can be reduced as low as desired.

Bennett Sample Pumps, Inc.

P.O. Box 7644

Amarillo, Texas 79114

Ph-806-352-0264

Fax-806-352-0264

website-www.bennettsamplepump.com